

Inhibin A ELISA

RUO

AL-123-r

INTENDED USE

The Inhibin A enzyme linked immunosorbent assay (ELISA) kit provides materials for the quantitative measurement of dimeric inhibin A in human serum. This assay is intended for research use only.

SUMMARY AND EXPLANATION

Inhibins are heterodimeric protein hormones secreted by granulosa cells of the ovary in the female and Sertoli cells of the testis in the male. They selectively suppress the secretion of pituitary follicle stimulating hormone (FSH) and have local paracrine actions in the gonads. The fully processed form of the inhibin molecule has a molecular weight of approximately 32 kDa and consists of the two distinct chains (α and β), linked by disulfide bridges. Higher molecular weight forms, with precursor forms of the α -subunit, also occur in follicular fluid and serum. In addition, free α -subunit forms, unassociated with a β -subunit, and lacking inhibin bioactivity, are also present. Inhibin A consists of an α -subunit and β_A -subunit. Measurements of Inhibin A are shown to be useful in studying its role in human reproductive physiology.

PRINCIPLE OF THE TEST

The Inhibin A ELISA is a quantitative three-step sandwich type immunoassay. In the first step Calibrators, Controls and unknown samples are added to Inhibin A antibody coated microtiter wells and incubated. After the first incubation and washing, the wells are incubated with biotinylated Inhibin A antibody. After the second incubation and washing, the wells are incubated with streptavidin horseradish peroxidase conjugate (SHRP). After the third incubation and washing step, the wells are incubated with substrate solution (TMB). After TMB incubation, an acidic stopping solution is added. In principle, the antibody-biotin conjugate binds to the solid phase antibody-antigen complex which in turn binds to the streptavidin-enzyme conjugate. The antibody-antigen-biotin conjugate-SHRP complex bound to the well is detected by enzyme-substrate reaction. The degree of enzymatic turnover of the substrate is determined by dual wavelength absorbance measurement at 450 nm as primary test filter and 630 nm as reference filter. The absorbance measured is directly proportional to the concentration of Inhibin A in the samples and calibrators.

MATERIALS SUPPLIED

CAL-123A Inhibin A Calibrators A (Lyophilized)

One vial, labeled A, containing 0 pg/mL Inhibin A in animal sera and a non-mercury preservative. Store unopened at 2-8°C until the expiration date. Reconstitute **calibrator A** with **1.0 mL** deionized water. Solubilize for **10 minutes**, mix well and use after reconstitution. Discard after 5 days, if stored at 2-8°C. For longer storage after reconstitution, aliquot and freeze at -20°C or colder for up to one year.

CAL-123B - CAL-123F Inhibin A Calibrators B thru F (Lyophilized)

Five vials, labeled B-F, containing concentrations of approximately 10-1200 pg/mL Inhibin A in animal sera and a non-mercury preservative. Refer to **calibration card** for exact concentrations. Store unopened at 2-8°C until the expiration date. Reconstitute calibrators B-F with **1.0 mL** deionized water. Solubilize for **10 minutes**, mix well and use after reconstitution. Discard after 5 days, if stored at 2-8°C. For longer storage after reconstitution, aliquot and freeze at -20°C or colder for up to one year.

STANDARDIZATION NOTE: The Inhibin A calibrators are traceable to the World Health Organization International preparation NIBSC code 91/624 version 3.0.

CTR-123-I & CTR-123-II Inhibin A Controls I & II (Lyophilized)

Two vials, labeled Levels I and II containing low and high Inhibin A concentrations in animal sera and a non-mercury preservative. Refer to **calibration card** for exact control ranges. Store unopened at 2-8°C until the expiration date. Reconstitute control Levels I and II with **1.0 mL** deionized water. Solubilize for **10 minutes**, mix well and use after reconstitution. Discard after 5 days, if stored at 2-8°C. For longer storage after reconstitution, aliquot and freeze at -20°C or colder for up to one year.

MI-123 Inhibin A Coated Microtitration Strips

One strip holder, containing 12 strips and 96 microtitration wells with Inhibin A antibody immobilized to the inside wall of each well. Store at 2-8°C until expiration date in the resealable pouch with a desiccant to protect from moisture.

ASB-123A Inhibin A Assay Buffer A

One bottle, 6 mL, containing a protein-based (BSA)-buffer with a non-mercury preservative. Store at 2-8°C until expiration date.

ASB-123B Inhibin A Assay Buffer B

One bottle, 6 mL, containing a buffer solution with a non-mercury preservative. Store at 2-8°C until expiration date.

BCC-123 Inhibin A Biotin Conjugate Concentrate

One vial, 0.4 mL, containing detection antibody biotin in a protein-based buffer with a non-mercury preservative. Dilute prior to use in Inhibin A Conjugate diluent. Store at 2-8°C until expiration date.

CND-123 Inhibin A Biotin Conjugate Diluent

One bottle, 12 mL, containing a protein-based buffer with a non-mercury preservative. Store at 2-8°C until expiration date.

SAR-123 Inhibin A Streptavidin-Enzyme Conjugate—Ready-to-Use (RTU)

One bottle, 12 mL, containing streptavidin-HRP (horseradish peroxidase) in a protein-based buffer and a non-mercury preservative. Store undiluted at 2-8°C until expiration date.

TMB-100 TMB Chromogen Solution

One bottle, 12 mL, containing a solution of tetramethylbenzidine (TMB) in buffer with hydrogen peroxide. Store at 2-8°C until expiration date.

STP-100 Stopping Solution

One bottle, 12 mL, containing 0.2 M sulfuric acid. Store at 2-30°C until expiration date.

WSH-100 Wash Concentrate A

One bottle, 60 mL, containing phosphate buffer saline solution with a nonionic detergent. Store at 2-30°C until expiration date. Dilute 25-fold with deionized water prior to use.

MATERIALS REQUIRED BUT NOT PROVIDED

1. Microplate reader capable of absorbance measurement at 450 nm, 405 nm and 630 nm.
2. Microplate orbital shaker.
3. Microplate washer.
4. Semi-automated/manual precision pipette to deliver 10–250 μL .
5. Vortex mixer.
6. Deionized water.

WARNINGS AND PRECAUTIONS

For *in vitro* research use only.

The following precautions should be observed:

- a) Follow good laboratory practice.
- b) Use personal protective equipment. Wear lab coats and disposable gloves when handling immunoassay materials.
- c) Handle and dispose of all reagents and material in compliance with applicable regulations.

WARNING: Potential Biohazardous Material

This reagent may contain some human source material (e.g. serum) or materials used in conjunction with human source materials. Handle all reagents and patient samples at a Biosafety Level 2, as recommended for any potentially infectious human material in the Centers for Disease Control/National Institutes of Health manual "Biosafety in Microbiological and Biomedical Laboratories," 6th Edition, 2020¹.

WARNING: Potential Chemical Hazard

Some reagents in this kit contain ProClin™ 300 and Sodium azide² as a preservative. ProClin™ 300 and Sodium azide in concentrated amounts are irritants to skin and mucous membranes.

For further information regarding hazardous substances in the kit, please refer to the MSDS, either at AnshLabs.com or by request.

SAMPLE COLLECTION AND PREPARATION

- a) Serum is the recommended sample type.
- b) Sample handling, processing, and storage requirements depend on the brand of blood collection tube that you use. Please reference the manufacturer's instructions for guidance. Each laboratory should determine the acceptability of its own blood collection tubes and serum separation products.
- c) Samples may be stored at 4°C if assayed within 24 hours; otherwise, samples must be stored at -20°C or -80°C to avoid loss of bioactivity and contamination.
- d) Avoid assaying lipemic, hemolyzed or icteric samples.
- e) Avoid repeated freezing and thawing of samples. Thaw samples no more than 3 times.
- f) For shipping, place specimens in leak proof containers in biohazard specimen bags with appropriate specimen identification and test

requisition information in the outside pocket of the biohazard specimen bag. Follow DOT and IATA requirements when shipping specimens.

PROCEDURAL NOTES

1. A thorough understanding of this package insert is necessary for successful use of the Inhibin A ELISA assay. It is the user's responsibility to validate the assay for their purpose. Accurate results will only be obtained by using precise laboratory techniques and following the package insert.
2. A calibration curve must be included with each assay.
3. Bring all kit reagents to room temperature ($23 \pm 2^\circ\text{C}$) before use. Thoroughly mix the reagents before use by gentle inversion. Do not mix various lots of any kit component and do not use any component beyond the expiration date.
4. Use a clean disposable pipette tip for each reagent, calibrator, control or sample. Avoid microbial contamination of reagents, contamination of the substrate solutions with the HRP conjugates. The enzyme used as the label is inactivated by oxygen, and is highly sensitive to microbial contamination, sodium azide, hypochlorous acid and aromatic chlorohydrocarbons often found in laboratory water supplies. Use deionized water.
5. Incomplete washing will adversely affect the outcome and assay precision. To minimize potential assay drift due to variation in the substrate incubation time, care should be taken to add the substrate solution into the wells. Avoid exposure of the reagents to excessive heat or direct sunlight during storage and incubation.

PREPARATION OF REAGENTS

1. **Inhibin A Calibrator A:** Tap and reconstitute Inhibin A Calibrator A with 1.0 mL deionized water. Solubilize for **10 minutes**, mix well and use after reconstitution.
2. **Inhibin A Calibrators B-F and Inhibin A Controls I & II:** Tap and reconstitute Inhibin A Calibrators B-F and Inhibin A Controls I & II each with 1.0 mL deionized water. Solubilize for **10 minutes**, mix well and use after reconstitution.
3. **Wash Solution:** Dilute wash concentrate 25-fold with deionized water. The wash solution is stable for one month at room temperature ($23 \pm 2^\circ\text{C}$) when stored in a tightly sealed bottle.
4. **Microtiter Wells:** Select the number of coated wells required for the assay. The remaining unused wells should be placed in the resealable pouch with a desiccant. The pouch must be resealed to protect from moisture.
5. **Inhibin A Antibody-Biotin Conjugate Solution:** The Inhibin A Antibody-Biotin Conjugate Concentrate should be diluted at a ratio of 1 part conjugate to 50 parts of Inhibin A Conjugate Diluent, according to the number of wells used. If an entire plate is to be used pipet exactly 220 μL of the Concentrate in to 11 mL of the diluent.
6. **Inhibin A Assay Buffers Premix Solution:** The Inhibin A Assay Buffer A (ASB-123A) and Inhibin A Assay Buffer B (ASB-123B) should be mixed by gentle inversion in equal volumes (1:1 ratio) according to the number of wells used. If an entire plate is to be used mix exactly 3 mL of the ASB-123 B in to 3 mL of the ASB-123 A. The pre-mixture solution is stable for use up to 4 hours. Discard the pre-mix solution after 4 hours.

ASSAY PROCEDURE

Allow all specimens and reagents to reach room temperature ($23 \pm 2^\circ\text{C}$) and mix thoroughly by gentle inversion before use. Calibrators, controls, and unknowns should be assayed in duplicate.

NOTE: All serum samples reading higher than the highest calibrator should be mixed and diluted in the 0 pg/mL reconstituted Calibrator A prior to assay.

1. Reconstitute Inhibin A Calibrators A-F with **1.0 mL** deionized water. Solubilize for **10 minutes**, Mix well.
2. Reconstitute Inhibin A Controls I & II each with **1.0 mL** deionized water. Solubilize for **10 minutes**, Mix well.
3. Label the microtitration strips to be used.
4. Pipette **50 µL** of the Calibrator, Controls and Unknowns to the appropriate wells.
5. Add **50 µL** of Inhibin A Assay Buffers Premix Solution (ASB-123A and ASB-123B in 1:1 ratio as described under the Preparation of the Reagents section of this package insert) to each well using a repeater pipette.
6. Incubate the plate, shaking at a fast speed (**600-800 rpm**) on an orbital microplate shaker, for **2.5 hour** at room temperature ($23 \pm 2^{\circ}\text{C}$).
7. During the last **20-30 minutes** of incubation, prepare the Inhibin A Antibody-Biotin Conjugate Solution by diluting the Inhibin A Biotin Conjugate Concentrate in Inhibin A Conjugate Diluent as described under the Preparation of the Reagents section of this package insert.
8. Aspirate and wash each strip **5 times** with Washing Solution (**350 µL/per well**) using an automatic microplate washer.
9. Add **100 µL** of the Antibody-Biotin Conjugate Solution to each well using a repeater pipette.
10. Incubate the plate, shaking at a fast speed (**600-800 rpm**) on an orbital microplate shaker, for **1 hour** at room temperature ($23 \pm 2^{\circ}\text{C}$).
11. Aspirate and wash each strip **5 times** with the Wash Solution (**350 µL/per well**) using an automatic microplate washer.
12. Add **100 µL** of the Streptavidin-Enzyme Conjugate-RTU to each well using a repeater pipette.
13. Incubate the plate, shaking at a fast speed (**600-800 rpm**) on an orbital microplate shaker, for **30 minutes** at room temperature ($23 \pm 2^{\circ}\text{C}$).
14. Aspirate and wash each strip **5 times** with the Wash Solution (**350 µL/per well**) using an automatic microplate washer.
15. Add **100 µL** of the TMB chromogen solution to each well using a repeater pipette. Avoid exposure to direct sunlight.
16. Incubate the wells, shaking at **600-800 rpm** on an orbital microplate shaker, for **10-12 min** at room temperature ($23 \pm 2^{\circ}\text{C}$).
NOTE: Visually monitor the color development to optimize the incubation time.
17. Add **100 µL** of the stopping solution to each well using a repeater pipette. Read the absorbance of the solution in the wells within **20 minutes**, using a microplate reader set to **450 nm**.
NOTE: Zero calibrator should be programmed as "Blank" while reading the optical density. If instrument has a wavelength correction, set the instrument to dual wavelength measurement at **450 nm** with background wavelength correction at **630 nm**.

RESULTS

NOTE: The results in this package insert were calculated by plotting the **log optical density (OD) data on the y-axis and log Inhibin A concentration on X-axis** using a cubic regression curve-fit. Alternatively, log vs. log quadratic regression curve-fit can be used. Other data reduction methods may give slightly different results.

1. Optimum results can be obtained at incubation temperature of $23 \pm 2^{\circ}\text{C}$.
2. Calculate the mean OD for each calibrator, Control, or Unknown.
3. Plot the log of the mean OD readings for each of the Calibrators along the y-axis versus log of the Inhibin A concentrations in pg/mL along the x-axis, using a cubic regression curve-fit.
4. Determine the Inhibin A concentrations of the Controls and unknowns from the calibration curve by matching their mean OD readings with the corresponding Inhibin A concentrations.
5. Any sample reading higher than the highest Calibrator should be appropriately diluted with the 0 pg/mL (CAL A) and re-assayed.

6. Any sample reading lower than the analytical sensitivity should be reported as such.
7. Multiply the value by a dilution factor, if required.

LIMITATIONS

The reagents supplied in this kit are optimized to measure Inhibin A levels in human serum. If there is evidence of microbial contamination or excessive turbidity in a reagent, discard the vial. For assays employing antibodies, the possibility exists for interference by heterophile antibodies in the samples⁴. The Inhibin A ELISA results should be interpreted with respect to the total clinical presentation of the patient, including: symptoms, clinical history, data from additional tests, and other appropriate patient examination information. The Inhibin A ELISA is not validated for use in prenatal screening to detect Down Syndrome, preeclampsia, ovarian tumors/cancers, and prediction of pregnancy in women undergoing IVF or to determine gonadal maturity in children.

QUALITY CONTROL

- Each laboratory should establish mean values and acceptable ranges to ensure proper performance.
- Inhibin A ELISA controls or other commercial controls should fall within established confidence limits.
- The confidence limits for Inhibin A controls are printed on the **Calibration card**.
- A full calibration curve, low- and high-level controls, should be included in each assay.
- TMB should be colorless. Development of any color may indicate reagent contamination or instability.

REPRESENTATIVE CALIBRATION CURVE DATA

Well Number	Well Contents	Mean OD	Conc (pg/mL)
Calibrators			
A1, A2	A	0.07 (Blank)	0
B1, B2	B	0.04	9.9
C1, C2	C	0.12	26.4
D1, D2	D	0.40	99
E1, E2	E	1.14	323.4
F1, F2	F	3.06	1188

CAUTION: The above data must not be employed in lieu of data obtained by the user in the laboratory.

ANALYTICAL CHARACTERISTICS

Analytical Sensitivity:

The analytical sensitivity in the assay as calculated by the interpolation of mean plus two standard deviation of 42 replicates in 21 runs of calibrator A (0 pg/mL) and calibrator B (9.9 pg/mL) is 5.45 pg/mL.

Imprecision:

Reproducibility of the Inhibin A assay was determined in a study using three serum pools. The study included a total of 21 assays, four replicates of each per assay (n=84). Representative data were calculated based on CLSI EP5-A2 guidelines and are presented in the following table.

Sample	Mean Conc.	Within Run		Between Run		Total	
		SD	%CV	SD	%CV	SD	%CV
Pool-1	101.34	4.8	4.7	4.2	4.1	6.3	6.3
Pool-2	344.85	11.6	3.4	14.8	4.3	18.9	5.5
Pool-3	31.83	2.0	6.1	1.8	5.6	2.6	8.3
Pool-4	84.29	4.7	5.6	3.4	4.0	5.8	6.9

Recovery:

Known amounts of Inhibin A were added to four serum samples containing different levels of endogenous Inhibin A. The concentration of Inhibin A was determined before and after the addition of exogenous Inhibin A and the percent recovery was calculated.

Sample	Endogenous Conc. (pg/mL)	Expected Conc. (pg/mL)	Observed Conc. (pg/mL)	% Recovery
1	132.1	182.387	180.146	99%
		228.096	209.307	92%
		269.831	256.453	95%
2	165.2	213.896	210.796	99%
		258.174	231.284	90%
		298.601	283.523	95%
3	187.1	234.728	236.096	101%
		278.058	270.861	97%
		317.621	292.182	92%

Linearity:

Based on CLSI EP6P-A multiple dilutions of the three serum samples containing various Inhibin A levels were diluted with calibrator A. The % recovery on individual samples is represented in the following table.

Sample	Dilution Factor	Expected Conc. (pg/mL)	Observed Conc. (pg/mL)	% Recovery
1	Neat	1188.0	N/A	N/A
	1:2	594.0	597.5	101
	1:4	297.0	281.4	95
	1:8	148.5	150.2	101
	1:16	74.2	65.9	89
	1:32	37.1	34.1	92
2	Neat	645.7	N/A	N/A
	1:2	322.9	340.5	105
	1:4	161.4	173.4	107
	1:8	80.7	86.8	108
	1:16	40.4	39.3	97
	3	Neat	212.150	N/A
1:2		106.1	114.3	108
1:4		53.0	60.5	114
1:8		26.5	26.5	100
1:16		13.3	15.1	114

Analytical Specificity:

This monoclonal antibody pair used in the assay detects Inhibin A. Other related molecules at the concentration in the table below did not show any significant cross-reaction. Specificity to other species has not been determined.

Sample	Cross-reactant	Concentration	% Cross-reactivity
1	Inhibin A	1000 pg/mL	100%
2	Activin A	50 ng/mL	ND
3	Activin B	50 ng/mL	ND
4	Activin AB	50 ng/mL	ND
5	Follistatin 288	50 ng/mL	ND
6	Follistatin 315	50 ng/mL	ND
7	Inhibin B	1000 ng/mL	ND
8	Alpha 2 Macroglobulin	50 ng/mL	ND
9	FSH	50 ng/mL	ND
10	Myostatin	50 ng/mL	ND

Expected Value:

Serum samples were analyzed using Ansh Inhibin A ELISA. The expected ranges for Inhibin A were calculated using 95% non-parametric estimation using Analyse-It® for Microsoft Excel.

Population	No of specimens (n)	Median conc. (pg/mL)	Inhibin A (pg/mL)
Normal Cycling Females			
Early follicular phase (-14 to -10)	16	13.0	5.3 - 22.2
Mid follicular phase (-9 to -4)	43	19.5	5.3 - 72.1
Late follicular phase (-3 to -1)	23	51.7	22 - 115
Mid cycle (Day 0)	7	99.4	90 - 151
Early Luteal (1 to 3)	25	73.5	29 - 209
Mid Luteal (4 to 11)	47	45.1	9 - 213
Late Luteal (12 to 14)	18	12.5	5.9 - 26
Post-Menopausal Females (Years)			
54-74	39	1.17	ND-5.24
Males Age (Years)			
3 - 10	9	4.53	2.3 - 6.5
11 - 24	35	2.9	1.0 - 10.3
25 - 40	26	3.41	2.0 - 5.5
41 - 60	34	3.4	2.1 - 8.1
61 - 77	68	2.68	1.0 - 6.8

Note: It is recommended that each laboratory should determine the reference range(s) for its own patient population. The results of this assay should be used in conjunction with other relevant and applicable clinical information.

Interference:

When hemoglobin, triglycerides and biotin were added at a concentration greater than two folds of their physiological concentration to control sample, average Inhibin A concentration were within $\pm 15\%$ of the control as represented in the following table.

Interferent	Interferent Dose	Sample Inhibin A (pg/mL)	Dosed Sample Inhibin A (pg/mL)	Inhibin A Difference (pg/mL)	% Difference to Reference
Hemoglobin	68 mg/dL	100.7	89.6	-11.2	-11.1
		125.5	116.0	-9.6	-7.6
		178.9	175.8	-3.0	-1.7
		149.3	143.3	-6.0	-4.0
		102.1	99.6	-2.5	-2.5
Triglycerides	500 mg/dL	128.5	130.6	2.0	1.6
		191.2	204.9	9.7	5.0
		149.8	149.5	5.7	4.0
		100.7	100.1	-0.6	-0.6
		125.5	128.0	2.4	2.0
Triglycerides	250 mg/dL	178.9	188.9	10.0	5.6
		149.3	150.8	1.5	1.0
		249.7	259.2	9.5	3.8
Biotin	1200 ng/mL	114.6	109.0	-5.6	-4.9
		173.3	268.1	-5.2	-1.9
Biotin	600 ng/mL	116.1	117.4	1.3	1.1
		272.2	272.5	0.3	0.1
Biotin	200 ng/mL	119.5	117.8	-1.7	-1.5


Method Comparison:

The Inhibin A ELISA has been compared to commercially available Inhibin A kit (Method A) using 116 pregnant female serum samples in the range of 50-795 pg/mL.

Passing Bablok analysis of the results yielded the following Regression:

Inhibin A ELISA (AL-123) = 1.03 (Method A) -14.9

($r=0.97$; $P<0.0001$)



REFERENCES

1. HHS Publication, 6th ed., 2020. Biosafety in Microbiological and Biomedical Laboratories. Available https://www.cdc.gov/labs/pdf/SF_19_203133-A_BMBL6_00-BOOK-WEB-final-3.pdf
2. DHHS (NIOSH) Publication No. 75-127, August 1976. Current Intelligence Bulletin 13 - Explosive Acids Hazard. Available <http://www.cdc.gov/niosh>.
3. Clinical and Laboratory Standards Institute (CLSI). Procedures for the Handling and Processing of Blood Specimens for Common Laboratory Tests; Approved Guideline—Fourth Edition. CLSI document H18-A4. Wayne, PA: Clinical and Laboratory Standards Institute 2010.
4. Kricka L. Interferences in immunoassays – still a threat. Clin Chem 2010; 46: 1037–1038.

This assay is intended for research use only.

The Ansh Labs logo is a trademark of Ansh Labs.



Manufactured by:
Ansh Labs
445 Medical Center Blvd.
Webster, TX 77598-4217, U.S.A.

Refer to illustrative Purposes Only.
the product for exact specifications.