

# Inhibin A ELISA



## AL-123-i

### INTENDED USE

The Inhibin A enzyme linked immunosorbent assay (ELISA) kit provides materials for the quantitative measurement of dimeric inhibin A in human serum. It is intended for *in vitro* diagnostic use as an aid in the diagnosis and monitoring of various hormonal reproductive disorders. Down syndrome (Trisomy 21) screening in the 2nd trimester using combined biochemical and ultrasound markers may be assessed with appropriate risk calculation software.<sup>1,2</sup> It is strongly recommended to use validated software (CE marked) designated specifically for evaluating the risk of Trisomy 21 using Inhibin A in combination with hAFP, hCG, and uE3 measurements.

### SUMMARY AND EXPLANATION

Inhibins are heterodimeric protein hormones secreted by granulosa cells of the ovary in the female and Sertoli cells of the testis in the male. They selectively suppress the secretion of pituitary follicle stimulating hormone (FSH) and also have local paracrine actions in the gonads. The fully processed form of the inhibin molecule has a molecular weight of approximately 32 kD and consists of the two distinct chains ( $\alpha$  and  $\beta$ ), linked by disulfide bridges. Higher molecular weight forms, with precursor forms of the  $\alpha$ -subunit, also occur in follicular fluid and serum. In addition, free  $\alpha$ -subunit forms, unassociated with a  $\beta$ -subunit, and lacking inhibin bioactivity, are also present. Inhibin A consists of an  $\alpha$ -subunit and  $\beta_A$ -subunit. Measurements of Inhibin A are shown to be useful in studying its role in human reproductive physiology. In uncomplicated pregnancy, the maternal serum levels of inhibin A first decrease, and then remain relatively constant in the first and second trimesters of pregnancy. In the second trimester of pregnancy, inhibin A concentrations are, on average, doubled in pregnancies affected by Down syndrome.<sup>3-5</sup> When screening is performed in the second trimester, upgrading from the Triple to the Quad test may well be the only option that allows national screening guidelines to be met. Inhibin A is also used in integrated and serum integrated screening strategies.

### PRINCIPLE OF THE TEST

The Inhibin A ELISA is a quantitative three-step sandwich type immunoassay. In the first step Calibrators, Controls and unknown samples are added to Inhibin A antibody coated microtiter wells and incubated. After the first incubation and washing, the wells are incubated with biotinylated Inhibin A antibody. After the second incubation and washing, the wells are incubated with streptavidin horseradish peroxidase conjugate (SHRP). After the third incubation and washing step, the wells are incubated with substrate solution (TMB). After TMB incubation, an acidic stopping solution is added. In principle, the antibody-biotin conjugate binds to the solid phase antibody-antigen complex which in turn binds to the streptavidin-enzyme conjugate. The antibody-antigen-biotin conjugate-SHRP complex bound to the well is detected by enzyme-substrate reaction. The degree of enzymatic turnover of the substrate is determined by dual wavelength absorbance measurement at 450 nm as primary test filter and 630 nm as reference filter. The absorbance measured is directly proportional to the concentration of Inhibin A in the samples and calibrators.

### MATERIALS SUPPLIED

#### CAL-123A Inhibin A Calibrator A (Lyophilized)

One vial, labeled A, containing 0 pg/mL Inhibin A in animal sera and a non-mercury preservative. Store unopened at 2-8°C until the expiration date. Reconstitute **calibrator A** with **1.0 mL** deionized water. Solubilize for **10 minutes**, mix well, and use after reconstitution. Discard after 5 days, if stored at 2-8°C. For longer storage after reconstitution, aliquot and freeze at -20°C or colder for up to one year.

#### CAL-123B - CAL-123F Inhibin A Calibrators B - F (Lyophilized)

Five vials, labeled B-F, containing concentrations of approximately 10-1200 pg/mL Inhibin A in animal sera and a non-mercury preservative. Refer to **calibration card** for exact concentrations. Store unopened at 2-8°C until the expiration date. Reconstitute calibrators B-F with **1.0 mL** deionized water. Solubilize for **10 minutes**, mix well, and use after reconstitution. Discard after 5 days, if stored at 2-8°C. For longer storage after reconstitution, aliquot and freeze at -20°C or colder for up to one year.

**STANDARDIZATION NOTE:** The Inhibin A calibrators are traceable to the World Health Organization International preparation NIBSC code 91/624 version 3.0.

#### CTR-123-I & CTR-123-II Inhibin A Controls I & II (Lyophilized)

Two vials, labeled Levels I and II containing low and high Inhibin A concentrations in animal sera and a non-mercury preservative. Refer to **calibration card** for exact control ranges. Store unopened at 2-8°C until the expiration date. Reconstitute control Levels I and II with **1.0 mL** deionized water. Solubilize for **10 minutes**, mix well, and use after reconstitution. Discard after 5 days, if stored at 2-8°C. For longer storage after reconstitution, aliquot and freeze at -20°C or colder for up to one year.

#### PLT-123 Inhibin A Coated Microtitration Strips

One strip holder, containing 12 strips and 96 microtitration wells with Inhibin A antibody immobilized to the inside wall of each well. Store at 2-8°C until expiration date in the resealable pouch with a desiccant to protect from moisture.

#### ASB-123A Inhibin A Assay Buffer A

One bottle, 6 mL, containing a protein-based (BSA)-buffer with a non-mercury preservative. Store at 2-8°C until expiration date.

#### ASB-123B Inhibin A Assay Buffer B

One bottle, 6 mL, containing a buffer solution with a non-mercury preservative. Store at 2-8°C until expiration date.

#### BCC-123 Inhibin A Biotin Conjugate Concentrate

One vial, 0.4 mL, containing detection antibody biotin in a protein-based buffer with a non-mercury preservative. Dilute prior to use in Inhibin A Conjugate diluent. Store at 2-8°C until expiration date.

**CND-123 Inhibin A Biotin Conjugate Diluent**

One bottle, 12 mL, containing a protein-based buffer with a non-mercury preservative. Store at 2-8°C until expiration date.

**SAR-123 Inhibin A Streptavidin-Enzyme Conjugate—Ready-to-Use (RTU)**

One bottle, 12 mL, containing streptavidin-HRP (horseradish peroxidase) in a protein-based buffer and a non-mercury preservative. Store undiluted at 2-8°C until expiration date.

**TMB-100 TMB Chromogen Solution**

One bottle, 12 mL, containing a solution of Tetramethylbenzidine (TMB) in buffer with hydrogen peroxide. Store at 2-8°C until expiration date.

**STP-100 Stopping Solution**

One bottle, 12 mL, containing 0.2 M sulfuric acid. Store at 2-30°C until expiration date.

**WSH-100 Wash Concentrate A**

One bottle, 60 mL, containing phosphate buffer saline solution with a nonionic detergent. Store at 2-30°C until expiration date. Dilute 25-fold with deionized water prior to use.

**MATERIALS REQUIRED BUT NOT PROVIDED**

1. Microplate reader capable of absorbance measurement at 450 nm, 405 nm and 630 nm.
2. Microplate orbital shaker.
3. Microplate washer.
4. Semi-automated/manual precision pipette to deliver 10–250  $\mu$ L.
5. Vortex mixer.
6. Deionized water.

**WARNINGS AND PRECAUTIONS****For *in vitro*-diagnostic use.**

The following precautions should be observed:

- a) Follow good laboratory practice.
- b) Use personal protective equipment. Wear lab coats and disposable gloves when handling immunoassay materials.
- c) Handle and dispose of all reagents and material in compliance with applicable regulations.
- d) If external package is damaged, inspect the components inside for any other damage. Do not use if the components are damaged.

**WARNING: Potential Biohazardous Material**

Samples and blood-derived products may be routinely processed with minimum risk using the procedure described. However, handle these products as potentially infectious according to universal precautions and good clinical laboratory practices, regardless of their origin, treatment, or prior certification.<sup>6</sup> Use an appropriate disinfectant for decontamination. Store and dispose of these materials and their containers in accordance with local regulations and guidelines.

**WARNING: Potential Chemical Hazard**

Some reagents in this kit contain Pro-Clean 400 and Sodium azide<sup>7</sup> as a preservative. Pro-Clean 400 and Sodium Azide in concentrated amounts are irritants to skin and mucous membranes.

For further information regarding hazardous substances in the kit, please refer to the MSDS, either at AnshLabs.com or by request.

**SAMPLE COLLECTION AND PREPARATION**

- a) Serum is the recommended sample type.
- b) Sample handling, processing, and storage requirements depend on the brand of blood collection tube that you use. Please reference the manufacturer's instructions for guidance. Each laboratory should

determine the acceptability of its own blood collection tubes and serum separation products.

- c) Samples may be stored at 4°C if assayed within 24 hours; otherwise, samples must be stored at -20°C or -80°C to avoid loss of bioactivity and contamination.
- d) Avoid assaying lipemic, hemolyzed or icteric samples.
- e) Avoid repeated freezing and thawing of samples. Thaw samples no more than 3 times.
- f) For shipping, place specimens in leak proof containers in biohazard specimen bags with appropriate specimen identification and test requisition information in the outside pocket of the biohazard specimen bag. Follow DOT and IATA requirements when shipping specimens<sup>8</sup>.

**PROCEDURAL NOTES**

1. A thorough understanding of this package insert is necessary for successful use of the Inhibin A ELISA assay. It is the user's responsibility to validate the assay for their purpose. Accurate results will only be obtained by using precise laboratory techniques and following the package insert.
2. A calibration curve must be included with each assay.
3. Bring all kit reagents to room temperature ( $23 \pm 2^\circ\text{C}$ ) before use. Thoroughly mix the reagents before use by gentle inversion. Do not mix various lots of any kit component and do not use any component beyond the expiration date.
4. Use a clean disposable pipette tip for each reagent, calibrator, control, or sample. Avoid microbial contamination of reagents, contamination of the substrate solutions with the HRP conjugates. The enzyme used as the label is inactivated by oxygen, and is highly sensitive to microbial contamination. Sodium Azide, Hypochlorous acid and aromatic Chlorohydrocarbons often found in laboratory water supplies. Use deionized water.
5. Incomplete washing will adversely affect the outcome and assay precision. To minimize potential assay drift due to variation in the substrate incubation time, care should be taken to add the substrate solution into the wells. Avoid exposure of the reagents to excessive heat or direct sunlight during storage and incubation.

**PREPARATION OF REAGENTS**

1. **Inhibin A Calibrators A-F and Inhibin A Controls I & II:** Tap and reconstitute Inhibin A Calibrators A-F and Inhibin A Controls I & II each with 1.0 mL deionized water. Solubilize for 10 minutes, mix well, and use after reconstitution.
2. **Wash Solution:** Dilute wash concentrate 25-fold with deionized water. The wash solution is stable for one month at room temperature ( $23 \pm 2^\circ\text{C}$ ) when stored in a tightly sealed bottle.
3. **Microtitration Wells:** Select the number of coated wells required for the assay. The remaining unused wells should be placed in the resealable pouch with a desiccant. The pouch must be resealed to protect from moisture.
4. **Inhibin A Antibody-Biotin Conjugate Solution:** The Inhibin A Antibody-Biotin Conjugate Concentrate should be diluted at a ratio of 1 part conjugate to 50 parts of Inhibin A Conjugate Diluent, according to the number of wells used. If an entire plate is to be used pipet exactly 220  $\mu$ L of the Concentrate in to 11 mL of the diluent.
5. **Inhibin A Assay Buffers Premix Solution:** The Inhibin A Assay Buffer A (ASB-123A) and Inhibin A Assay Buffer B (ASB-123B) should be mixed by gentle inversion in equal volumes (1:1 ratio) according to the number of wells used. If an entire plate is to be used mix exactly 3 mL of the ASB-123 B in to 3 mL of the ASB-123 A. The pre-mixture solution is stable for use up to 4 hours. Discard the pre-mix solution after 4 hours.

## ASSAY PROCEDURE

Allow all specimens and reagents to reach room temperature ( $23 \pm 2^\circ\text{C}$ ) and mix thoroughly by gentle inversion before use. Calibrators, controls, and unknowns should be assayed in duplicate.

**NOTE:** All serum samples reading higher than the highest calibrator should be mixed and diluted in the 0 pg/mL reconstituted Calibrator A prior to assay.

1. Reconstitute Inhibin A Calibrators A-F and Inhibin A Controls I & II each with **1.0 mL** deionized water. Solubilize for **10 minutes** and mix well.
2. Label the microtitration strips to be used.
3. Pipette **50  $\mu\text{L}$**  of the Calibrator, Controls and Unknowns to the appropriate wells.
4. Add **50  $\mu\text{L}$**  of Inhibin A Assay Buffers Premix Solution (ASB-123A and ASB-123B in 1:1 ratio as described under the Preparation of the Reagents section of this package insert) to each well using a repeater pipette.
5. Incubate the plate, shaking at a fast speed (**600-800 rpm**) on an orbital microplate shaker, for **2.5 hour** at room temperature ( $23 \pm 2^\circ\text{C}$ ).
6. During the last **20-30 minutes** of incubation, prepare the Inhibin A Antibody-Biotin Conjugate Solution by diluting the Inhibin A Biotin Conjugate Concentrate in Inhibin A Conjugate Diluent as described under the Preparation of the Reagents section of this package insert.
7. Aspirate and wash each strip **5 times** with Washing Solution (**350  $\mu\text{L}$ /per well**) using an automatic microplate washer.
8. Add **100  $\mu\text{L}$**  of the Antibody-Biotin Conjugate Solution to each well using a repeater pipette.
9. Incubate the plate, shaking at a fast speed (**600-800 rpm**) on an orbital microplate shaker, for **1 hour** at room temperature ( $23 \pm 2^\circ\text{C}$ ).
10. Aspirate and wash each strip **5 times** with the Wash Solution (**350  $\mu\text{L}$ /per well**) using an automatic microplate washer.
11. Add **100  $\mu\text{L}$**  of the Streptavidin-Enzyme Conjugate-RTU to each well using a repeater pipette.
12. Incubate the plate, shaking at a fast speed (**600-800 rpm**) on an orbital microplate shaker, for **30 minutes** at room temperature ( $23 \pm 2^\circ\text{C}$ ).
13. Aspirate and wash each strip **5 times** with the Wash Solution (**350  $\mu\text{L}$ /per well**) using an automatic microplate washer.
14. Add **100  $\mu\text{L}$**  of the TMB chromogen solution to each well using a repeater pipette. Avoid exposure to direct sunlight.
15. Incubate the wells, shaking at **600-800 rpm** on an orbital microplate shaker, for **10-12 min** at room temperature ( $23 \pm 2^\circ\text{C}$ ).
16. Add **100  $\mu\text{L}$**  of the stopping solution to each well using a repeater pipette. Read the absorbance of the solution in the wells within **20 minutes**, using a microplate reader set to **450 nm**.

**NOTE:** Visually monitor the color development to optimize the incubation time.

**NOTE:** Zero calibrator should be programmed as "Blank" while reading the optical density. If instrument has a wavelength correction, set the instrument to dual wavelength measurement at **450 nm** with background wavelength correction at **630 nm**.

## RESULTS

**NOTE:** The results in this package insert were calculated by plotting the log optical density (OD) data on the y-axis and log Inhibin A concentration on X-axis using a cubic regression curve-fit. Alternatively, log vs. log quadratic regression curve-fit can be used. Other data reduction methods may give slightly different results.

1. Optimum results can be obtained at incubation temperature of  **$23 \pm 2^\circ\text{C}$** .
2. Calculate the mean OD for each calibrator, Control, or Unknown.
3. Plot the log of the mean OD readings for each of the Calibrators along the y-axis versus log of the Inhibin A concentrations in pg/mL along the x-axis, using a cubic regression curve-fit.

4. Determine the Inhibin A concentrations of the Controls and unknowns from the calibration curve by matching their mean OD readings with the corresponding Inhibin A concentrations.
5. Any sample reading higher than the highest Calibrator should be appropriately diluted with the 0 pg/mL (CAL A) and re-assayed.
6. Any sample reading lower than the analytical sensitivity should be reported as such.
7. Multiply the value by a dilution factor, if required.

## LIMITATIONS

The reagents supplied in this kit are optimized to measure Inhibin A levels in human serum. If there is evidence of microbial contamination or excessive turbidity in a reagent, discard the vial. For assays employing antibodies, the possibility exists for interference by heterophile antibodies in the samples<sup>9</sup>.

The Inhibin A ELISA results should be interpreted with respect to the total clinical presentation of the patient, including: symptoms, clinical history, data from additional tests, and other appropriate patient examination information.

## QUALITY CONTROL

- Each laboratory should establish mean values and acceptable ranges to assure proper performance.
- Inhibin A ELISA controls or other commercial controls should fall within established confidence limits.
- The confidence limits for Inhibin A controls are printed on the **Calibration card**.
- A full calibration curve, low- and high-level controls, should be included in each assay.
- TMB should be colorless. Development of any color may indicate reagent contamination or instability.

## REPRESENTATIVE CALIBRATION CURVE DATA

Well Number	Well Contents	Mean OD	Conc (pg/mL)
A1, A2	Calibrators A	0.07 (Blank)	0
B1, B2	B	0.04	9.9
C1, C2	C	0.12	26.4
D1, D2	D	0.40	99.0
E1, E2	E	1.14	323.4
F1, F2	F	3.06	1188.0

**CAUTION:** The above data must not be employed in lieu of data obtained by the user in the laboratory.

## ANALYTICAL CHARACTERISTICS

### Analytical Sensitivity:

The analytical sensitivity in the assay as calculated by the interpolation of mean plus two standard deviation of 42 replicates in 21 runs of calibrator A (0 pg/mL) and calibrator B (9.9 pg/mL) is 5.45 pg/mL.

### Imprecision:

Reproducibility of the Inhibin A assay was determined in a study using Four serum pools. The study included a total of 21 assays, four replicates of each per assay (n=84). Representative data calculated are presented in the following table.

Sample	Mean Conc.	Within Run		Between Run		Total	
		(pg/mL)	SD	%CV	SD	%CV	SD
Pool-1	101.34	4.8	4.7	4.2	4.1	6.3	6.2
Pool-2	344.85	11.6	3.4	14.8	4.3	18.8	5.5
Pool-3	31.83	1.9	6.1	1.8	5.6	2.6	8.3
Pool-4	84.293	4.7	5.6	3.4	4.0	5.8	6.9

#### Recovery:

Known amounts of Inhibin A were added to five serum samples containing different levels of endogenous Inhibin A. The concentration of Inhibin A was determined before and after the addition of exogenous Inhibin A and the percent recovery was calculated.

Sample ID	Endogenous Value in pg/mL	Expected in pg/mL	Observed in pg/mL	%Recovery	Average %Recovery
1	68.9	122.2	109.8	90%	85%
		170.6	138.8	81%	
		214.9	179.3	83%	
2	80.6	133.3	127.1	95%	89%
		181.3	159.1	88%	
		225.0	187.4	83%	
3	132.1	182.4	180.1	99%	95%
		228.1	209.3	92%	
		269.8	256.5	95%	
4	187.1	234.7	236.1	101%	97%
		278.1	270.9	97%	
		317.6	292.2	92%	
5	165.2	213.9	210.8	99%	94%
		258.2	231.3	90%	
		298.6	283.5	95%	

#### Linearity:

Based on NCCLS EP-6-P, multiple dilutions of the four serum samples and Calibrator F containing various Inhibin A levels were diluted with calibrator A.

The % recovery on individual samples is represented in the following table

Sample ID	Dilution factor (1 in X)	Expected Value in pg/mL	Observed Value in pg/mL	%Recovery	Average %Recovery
Calibrator F	Neat	1188.0			93%
	1:2	594.0	597.5	101%	
	1:4	297.0	281.4	95%	
	1:8	148.5	130.2	88%	
	1:16	74.3	65.9	89%	
	1:32	37.1	34.1	92%	
1	Neat	364.8			109%
	2	182.4	213.6	117%	
	4	91.2	102.3	112%	
	8	45.6	48.9	107%	
	16	22.8	22.7	100%	
2	Neat	212.2			109%
	2	106.1	114.3	108%	
	4	53.0	60.5	114%	
	8	26.5	26.5	100%	
3	Neat	645.7			104%
	2	322.8	340.5	105%	
	4	161.4	173.4	107%	
	8	80.7	86.8	108%	
4	Neat	418.2			93%
	2	209.1	216.6	104%	
	4	104.5	101.3	97%	
	8	52.3	45.3	87%	
	16	26.1	22.2	85%	

#### Analytical Specificity:

This monoclonal antibody pair used in the assay detects Inhibin A. Other related molecules at the concentration in the table below did not show any

significant cross-reaction. Specificity to other species has not been determined.

Sample	Cross-reactant	Concentration	% Cross-reactivity
1	Inhibin A	1000 pg/mL	100%
2	Activin A	50 ng/mL	ND
3	Activin B	50 ng/mL	ND
4	Activin AB	50 ng/mL	ND
5	Follistatin 288	50 ng/mL	ND
6	Follistatin 315	50 ng/mL	ND
7	Inhibin B	50 ng/mL	ND
8	Alpha 2 Macroglobulin	50 ng/mL	ND
9	FSH	50 ng/mL	ND
10	Myostatin	50 ng/mL	ND

#### Interference:

When hemoglobin, triglycerides and biotin were added at a concentration greater than two folds of their physiological concentration to control sample, average Inhibin A concentration were within  $\pm 15\%$  of the control as represented in the following table.

Interferent	Interferent Dose	Sample Inhibin A (pg/mL)	Dosed Sample Inhibin A (pg/mL)	Inhibin A Difference (pg/mL)	% Difference to Reference
Hemoglobin	68 mg/dL	100.7	89.6	-11.2	-11.1
		125.5	115.9	-9.6	-7.6
		178.8	175.8	-3.0	-1.7
		149.3	143.3	-6.0	-4.0
Triglycerides	500 mg/dL	102.1	99.6	-2.5	-2.5
		128.5	130.6	2.0	1.6
		195.2	204.9	9.7	5.0
		143.8	149.5	5.7	4.0
Triglycerides	250 mg/dL	100.7	100.1	-0.6	-0.6
		125.5	128.0	2.4	1.9
		178.8	188.9	10.0	5.6
		149.3	150.8	1.5	1.0
Biotin	1200 ng/mL	249.7	259.2	9.5	3.8
		114.6	109.0	-5.6	-4.9
Biotin	600 ng/mL	273.3	268.1	-5.2	-1.9
		116.1	117.4	1.3	1.1
Biotin	200 ng/mL	272.2	272.5	0.3	0.1
		119.5	117.8	-1.7	-1.5


#### Expected Value:

Serum samples were analyzed using Ansh Inhibin A ELISA. The expected ranges for Inhibin A were calculated using 95% non-parametric estimation using Analyse-It® for Microsoft Excel.

Population	No of specimens (n)	Median conc. (pg/mL)	Inhibin A (pg/mL)
<b>Normal Cycling Females</b>			
Early follicular phase (-14 to -10)	16	13.0	5.3 - 22.2
Mid follicular phase (-9 to -4)	43	19.5	5.3 - 72.1
Late follicular phase (-3 to -1)	23	51.7	22 - 115
Mid cycle (Day 0)	7	99.4	90 - 151
Early Luteal (1 to 3)	25	73.5	29 - 209
Mid Luteal (4 to 11)	47	45.1	9 - 213
Late Luteal (12 to 14)	18	12.5	5.9 - 26
<b>Post-Menopausal Females (Years)</b>			
54-74	39	1.17	ND - 5.24
<b>Males Age (Years)</b>			
3 - 10	9	4.53	2.3 - 6.5
11 - 24	35	2.9	1.0 - 10.3
25 - 40	26	3.41	2.0 - 5.5
41 - 60	34	3.4	2.1 - 8.1
61 - 77	68	2.68	1.0 - 6.8

The values represented in the graph and the table below are maternal serum inhibin A in the second trimester.

**Change in Inhibin A Concentration with Gestational Age**




Population	No of specimens (n)	Median conc. (pg/mL)	Inhibin A (pg/mL) 95% CI
<b>Gestational Age (weeks)</b>			
15	228	190.02	162.80 to 208.20
16	259	168.20	154.97 to 187.21
17	158	161.95	140.45 to 178.40
18	108	169.90	149.76 to 194.20
19	48	172.85	124.60 to 213.90
20	26	191.50	139.20 to 232.60
21	12	212.30	135.10 to 263.60
22	9	153.08	106.00 to 357.00

**Note:** It is recommended that each laboratory should determine the reference range(s) for its own patient population. The results of this assay should be used in conjunction with other relevant and applicable clinical information.

**Intra-Cycle Variability of E2, Inhibin B and Inhibin A**


The mean ± SE concentrations of inhibin A, estradiol, and Inhibin B in 9 subjects are shown, aligned relative to the day of the mid cycle LH peak. The inhibin A levels remained low (13 pg/mL) during the early follicular phase and rose in the late follicular phase to reach a peak concentration (99.4 mL) on the day of the LH peak. After the LH peak, inhibin A levels fell briefly, then rose again during early to mid-luteal phase, and fell again in concert with the drop in progesterone during the late luteal phase (12.5 pg/mL). Estradiol and inhibin A were highly correlated during the follicular phase (day -14 to -2) as evident by the following linear regression data:




**Method Comparison:**

The Inhibin A ELISA has been compared to commercially available Inhibin A kit (Method A) using 116 pregnant female serum samples in the range of 50-795 pg/mL.

Passing Bablok analysis of the results yielded the following Regression: Inhibin A ELISA (AL-123i) = 1.03 (Method A) -14.9 (r=0.97; P<0.0001)



Receiver Operating Characteristic (ROC): The Inhibin A ELISA has been compared to commercially available Inhibin A kit (Dxl) using a case-control study set of 45 cases (Down Syndrome) and 221 controls. A receiver operating characteristic (ROC) curve including Down syndrome samples indicated that the screening efficiency using the Ansh Inhibin A ELISA (AL123i) is comparable.




There is no statistically significant difference between these assays in the inhibin distribution parameters or the proportion of cases with elevated levels.

**REFERENCES**

1. Maymon R and Shulman A. Integrated first- and second trimester Down syndrome screening test among unaffected IVF pregnancies. *Prenat. Diagn.* 2004; 24: 125-129.
2. Wald NJ, Bestwick JP, Morris JK. Cross-trimester marker ratios in prenatal screening for Down syndrome. *Prenat. Diagn.* 2006; 26: 514-523
3. Wald NJ, Rodeck C, Hackshaw AK, Walters J, Chitty L, Mackinson AM. First and second trimester antenatal screening for Down's syndrome: the results of the Serum, Urine and Ultrasound Screening Study (SURUSS). *J Med Screen.* 2003; 10: 56-104.
4. Canick JA, MacRae A. Second trimester serum markers. *Semin Perinatol.* 2005; 29: 203-208.
5. Malone FD, Canick JA, Ball RH, et al. First- and second- trimester evaluation of risk (FASTER) research consortium. First-trimester or secondtrimester screening, or both, for Down syndrome. *N Engl J Med.* 2005; 353: 2001-2011.
6. HHS Publication, 5th ed., 2009. Biosafety in Microbiological and Biomedical Laboratories. Available: <http://www.cdc.gov/biosafety/publications/bmb15/BMBL>
7. DHHS (NIOSH) Publication No. 78-127, August 1976. Current Intelligence Bulletin 13 - Explosive Azide Hazard. Available: <http://www.cdc.gov/niosh>.
8. Approved Guideline – Procedures for the Handling and Processing of Blood Specimens, H18-A3. 2004. Clinical and Laboratory Standards Institute.

- 9. Kricka L. Interferences in immunoassays – still a threat. Clin Chem 2000; 46: 1037–1038.



This assay is intended for *in vitro* diagnostic use.



The Ansh Labs logo is a trademark of Ansh Labs.



Manufactured by:

Ansh Labs

445 Medical Center Blvd.

Webster, TX 77598-4217, U.S.A.



**Notified Body:**

*BSI Group The Netherlands B.V.*

*Say Building*

*John M. Keynesplein 9*

*1066 EP Amsterdam*

*Netherlands*

**European Representative:**

*RD-RatioDiagnostics GmbH*

*Westerbachstr.47*

*60489 Frankfurt*

*Germany*



Ansh Labs consumables are being shipped with English Instructions for Use (IFUs). You may contact your local Ansh Labs sales representative or technical support organization to obtain translated IFUs.

Les consommables pour Ansh Labs sont livrés avec des instructions d'utilisation en anglais. N'hésitez pas à contacter votre société d'assistance technique ou votre représentant Ansh Labs local pour obtenir des instructions traduites.


Die Verbrauchsmaterialien von Ansh Labs werden mit englischer Gebrauchsanweisung (IFU) geliefert. Wenden Sie sich gegebenenfalls an Ihren zuständigen Vertreter von Ansh Labs oder den technischen Kundendienst, um übersetzte Gebrauchsanweisungen zu erhalten.

Los consumibles para Ansh Labs se entregan con las instrucciones de uso (IFU) en inglés. También puede ponerse en contacto con el representante local de ventas de Ansh Labs o con la empresa de asistencia técnica para obtener las IFU traducidas.

European CE Mark Version  
For Illustrative Purposes Only  
The package insert included with  
the product for exact specifications.

**FOR EXPORT ONLY**

<b>Symbols used with Ansh Labs Assays</b>
---

Symbol	English	Deutsch	Français	Español	Italiano
	European Conformity	CE-Konformitätskennzeichnung	Conformité aux normes européennes	Conformidad europea	Conformità europea
	Biohazard	Biogefahr	danger biologique	Riesgo biológico	rischio biologico
	Caution	Vorsicht	mise en garde	precaución	attenzione
	Consult instructions for use	Gebrauchsanweisung beachten	Consulter les instructions d'utilisation	Consulte las instrucciones de uso	Consultare le istruzioni per l'uso
	In vitro diagnostic device	In-vitro-Diagnostikum	Usage Diagnostic in vitro	Para uso Diagnóstico in vitro	Per uso Diagnostica in vitro
	Catalogue number	Katalog-Nr.	Numéro de catalogue	Número de catálogo	Numero di Catalogo
	Lot. No. / Batch code	Chargen-Nr.	Numéro de lot	Número de lote	Numero di lotto
	Storage Temperature	Lagerungstemperatur	Température de conservation	Temperatura de conservación	Temperatura di conservazione
	Expiration Date	Mindesthaltbarkeitsdatum	Date limite d'utilisation	Fecha de caducidad	Data di scadenza
	Date of Manufacture	Herstellungsdatum	date de fabrication	fecha de manufactura	Data di produzione
	Legal Manufacturer	Hersteller	fabricant	Fabricante	Fabbricante
Content	Content	Inhalt	Conditionnement	Contenido	Contenuto
Volume/No.	Volume / No.	Volumen/Anzahl	Volume/Quantité	Volumen/Número	Volume/Quantità
	96-well plate	Platte mit 96 Vertiefungen	Plaque à 96 puits	placa de 96 pocillos	piastra a 96 pozzetti

European CE Mark Version  
 For illustrative purposes only  
 Refer to package insert included with  
 the product for exact specifications.